



## EC Declaration of conformity

This is to certify that the products

**CYCLE 400**  
**CYCLE 450**

fulfil the requirements laid down in the following guidelines and standards

- **2004/108/EC**  
EMC (electromagnetic compatibility) Directive
- **2006/95/EG**  
Low voltage directive
- **EN 957 HA**  
Stationary training equipment
- **DIN EN 60335-1, protection class II**  
Household and similar electrical appliances - Safety – Part 1: General requirements

This declaration is valid for all products delivered from 01.01.2012 until 31.12.2012 and has been issued with full responsibility to the producer

**ERGO-FIT GmbH & Co.KG**  
**Blocksbergstraße 165**  
**D-66955 Pirmasens**

Name of the CE document commissioner: **Alexander Harrer**  
Address of the CE document commissioner: **see manufacturer's address**

by:

A handwritten signature in black ink, appearing to read 'Michael Resch', written in a cursive style.

Michael Resch  
(Managing Director)

Pirmasens, 01.12.2011

Erstellt	Geprüft	Freigegeben	Dokument-Nr.
Am: 16.11.2011 Von: SPA	Am: 14.12.2011 Von: AHA	Am: 14.12.2011 Von: AHA	ERGO- FIT_Konformitaetserklaerung 2012-1_CL 400-Englisch- 20111219.doc19

## EC Declaration of conformity

This is to certify that the products  
**CYCLE 407 MED**  
**CYCLE 457 MED**

fulfil the requirements laid down in the following guidelines and standards.

- **MPG, MDD 93/42/EEC, class of models IIa, annex II**  
Medical Device Directive
- **EN 957 SA**  
Stationary training equipment
- **DIN EN 60601-1, class of protection I, degree of protection type B**  
Medical electrical equipment – Part 1: General requirements for safety and essential performance
- **DIN EN 60601-1-2**  
Medical electrical equipment – Part 1: General requirements for safety and essential performance –  
Collateral standard: Electromagnetic compatibility - Requirements and tests
- **DIN EN 60601-1-6**  
Medical electrical equipment – Part 1: General requirements for safety and essential performance –  
Collateral standard: Usability
- **DIN VDE 0750-238**  
Medical electrical equipment – Part 238 Particular requirements for the safety of crank ergometers
- **DIN EN 62304**  
Medical device software - Software life-cycle processes

Notified body:

**DQS GmbH**  
**August-Schanz-Straße 21**  
**D-60433 Frankfurt am Main**  
**Kennnummer: 0297**

This declaration is valid for all products delivered from 01.01.2012 until 31.12.2012 and has been issued with full responsibility to the producer

**ERGO-FIT GmbH & Co.KG**  
**Blocksbergstraße 165**  
**D-66955 Pirmasens**

Name of the CE document commissioner: **Alexander Harrer**  
Address of the CE document commissioner: **see manufacturer's address**

by:



Michael Resch  
(Managing Director)

Pirmasens, 01.12.2011

Erstellt	Geprüft	Freigegeben	Dokument-Nr.
Am: 16.11.2011 Von: SPA	Am: 14.12.2011 Von: AHA	Am: 14.12.2011 Von: AHA	ERGO- FIT_Konformitaetserklaerung 2012-1_CL 400-Englisch- 20111219.doc19